

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Particle Analyzer

Name and address of the applicant

Micromeritics Instrument Corporation
4356 Communications Dr.
Norcross, GA, 30093-2901
United States of America

Name and address of the manufacturer

Micromeritics Instrument Corporation
4356 Communications Dr.
Norcross, GA, 30093-2901
United States of America

Name and address of the factory

Note: When more than one factory, please report on page 2

Micromeritics Instrument Corporation
4356 Communications Dr.
Norcross, GA, 30093-2901
United States of America

☐ Additional information on page 2

Ratings and principal characteristics

100-240Vac, 50-60Hz, 625VA

Trademark / Brand (if any)



mi | micromeritics

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

Accusorp HP-2520

Additional information (if necessary may also be reported on page 2)

☐ Additional information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010, IEC 61010-1:2010/AMD1:2016, IEC 61010-2-010:2019,
IEC 61010-2-081:2019

Comments:

Refer to the Test Report for the Summary of Compliance with National
Deviations

As shown in the Test Report Ref. No. which forms part of this Certificate

106558563ATL-001C, 106558563ATL-001D, 106558563ATL-001E

This CB Test Certificate is issued by the National Certification Body

Intertek Testing Services NA, Inc.
545 E. Algonquin Road, Arlington Heights
IL 60005
United States of America

Date: 2026-09-01

Signature: Matt Snyder